

This product contains methane, oxygen and nitrogen, substances subject to the Pennsylvania Worker and Community Right-To-Know Act.

PRODUCT IDENTITY

LABEL IDENTITY - MSA P/N 463464 Calibration Gas, 2.5% Methane in Air
CHEMICAL NAME - Methane, Oxygen, Nitrogen Mixture
ADDITIONAL IDENTITIES - MSA P/N 463464 Calibration gas
FORMULA - CH₄ in Air

APPLICABLE CHEMICAL CONTENTS

| | <u>%</u> | <u>TWA</u> |
|-----------------------|----------|------------|
| Methane (CAS 74-82-8) | 2.5 | None* |
| Air | Balance | None |

*Methane is a simple asphyxiant (ACGIH 2008)

NOTE: Gas Under Pressure, 115 PSIG at 70°F, Approx. 3 Liters Gas at Atmospheric Pressure

PHYSICAL AND CHEMICAL PROPERTIES

APPEARANCE AND ODOR - Colorless, Odorless Gas
BOILING POINT - N/A
VAPOR PRESSURE - N/A
VAPOR DENSITY (AIR = 1) - < 1
SOLUBILITY IN WATER - Methane - 9 cm³/100 ml (20°C)
Oxygen - 3.2 cm³/100 ml (25°C)
Nitrogen - 2.3 cm³/100 ml (0°C)
SPECIFIC GRAVITY (H₂O = 1) - N/A
PERCENT VOLATILE BY VOLUME - N/A

N/A - Not Applicable

PHYSICAL HAZARD INFORMATION

PHYSICAL HAZARD - Compressed Gas, 115 PSIG at 70°F
CONDITIONS OR MATERIALS TO AVOID - None
FLASH POINT - N/A
LEL - N/A
UEL - N/A
EXTINGUISHING MEDIA - This Calibration Gas Mixture is not flammable
SPECIAL FIRE FIGHTING PROCEDURES - See Next Item
UNUSUAL FIRE AND EXPLOSION HAZARDS - Gas Under Pressure, 115 PSIG at 70°F. Do Not Exceed 120°F.

HEALTH HAZARDS

HEALTH HAZARDS - Methane is a simple asphyxiant, ACGIH 2003.

SIGNS AND SYMPTOMS OF EXPOSURE - None known for 2.5% Methane in air.

PRIMARY ROUTES OF ENTRY - Inhalation

TARGET ORGANS - Lungs

MEDICAL CONDITIONS GENERALLY RECOGNIZED AS BEING AGGRAVATED BY EXPOSURE - No Information

EXPOSURE LIMITS - None. Methane is a simple asphyxiant (ACGIH 2008.)

CARCINOGENICITY DATA - NIOSH RTECS, OSHA, NTP or IARC does not list component gases.

EMERGENCY AND FIRST AID PROCEDURES - None

SAFE HANDLING AND USE

HYGIENIC PRACTICES - Avoid breathing gas.

PROTECTIVE MEASURES DURING REPAIR AND MAINTENANCE OF CONTAMINATED EQUIPMENT - N/A

PROCEDURES FOR SPILL OR LEAK CLEANUP - Ventilate area

WASTE DISPOSAL - Do not puncture or incinerate cylinder. Before discarding cylinder, slowly release contents to a safe exhaust.

STORAGE - Store in a cool, dry, well-ventilated area. Do not exceed 120°F.

CONTROL MEASURES

PERSONAL PROTECTIVE EQUIPMENT - Due to the limited amount of gas in the cylinder, and the low release rate employed in instrument calibration, respiratory protection is not indicated under conditions of intended use.

ENGINEERING CONTROLS - Mechanical ventilation is suitable.

WORK PRACTICES - Avoid breathing gas. Use in well-ventilated areas. Follow the calibration procedure detailed in the MSA instruction manual provided with the instrument under calibration.

DATE OF PREPARATION - Rev. 11, January 2009

WARNING: This is a hazardous chemical product. By following the directions and warnings provided with this product, the hazards associated with the use of this product can be greatly reduced but never entirely eliminated. Mine Safety Appliances Company makes no warranties, expressed or implied, with respect to this product and EXPRESSLY DISCLAIMS THE WARRANTY OF MERCHANTABILITY AND ANY WARRANTY OF FITNESS FOR A PARTICULAR PURPOSE. Users assume all risks in handling, using or storing this product.