



**VERIFY™ Chemical Indicator for SYSTEM 1E® Liquid
Chemical Sterilant Processing System**
Safety Data Sheet

Version: 1.0

1. Identification of the Substance and Company

**VERIFY™ Chemical Indicator for SYSTEM 1E® Liquid Chemical
Sterilant Processing System**

Product No. LCC016
SDS No. S288US

NFPA 704 HAZARD RATING:

HEALTH: ND
FIRE: ND
REACTIVITY: ND

Prepared by: M. Ebers
asksteris_msds@steris.com

Date Created: September 13, 2010

Date Revised: August 15, 2014

Date Reviewed: NA

STERIS Corporation, P.O. Box 147, St. Louis, Mo 63166, US
Emergency Telephone No. 1-314-535-1395 (STERIS); 1-800-424-9300 (CHEMTREC)
Telephone Number for information: 1-800-548-4873 (Customer Service – Healthcare Products)

Albert Browne Limited, Chancery House, 190 Waterside Road, Hamilton Industrial Park,
Leicester, LE5 1QZ, UK
Emergency Phone No: +44 (0) 1895 622 639
Product/Technical Information Phone No: +44 (0) 116 276 8636

PRODUCT NAME:	PRODUCT NUMBER:
VERIFY™ Chemical Indicator for SYSTEM 1E® Liquid Chemical Sterilant Processing System	LCC016
SHIPPING INFORMATION:	NOT RESTRICTED

This product is NOT a hazardous product according to the OSHA Hazard Communication Standard, 29 CFR 1910.1200.

VERIFY™ Chemical Indicator for System 1E® Liquid Chemical Sterilant Processing System does not contain any components derived from animal origin, does not come into contact with materials of animal origin and therefore complies with the "Note for guidance of minimizing the risk of transmitting animal spongiform encephalopathy agents via human Veterinary medicinal products (EMEA/410/01 Rev. 2-October 2003)."

In The Case Of Accidental Exposure, Take The Following Precautionary Measures:

EYES: Immediately flush eyes with plenty of water for 15 minutes. Get medical attention if irritation persists.

SKIN: Immediately wash with soap and water. Get medical attention if irritation develops.

INGESTION: Give water to drink and get medical advice.

INHALATION: Remove to fresh air. Get medical attention for any breathing difficulty.

Chemical Indicators Handling And Disposal:

Handling and Storage Precautions: Keep cap tightly closed. Store at 6°C (43°F) to 30°C (86°F) and 30-60% relative humidity.

Waste Disposal Methods: Dispose of used/expired test strips and plastic bottles in trash receptacle in accordance with federal, state, and local regulations.

The information on this sheet is not a specification and does not guarantee specific properties. The information is intended to provide general knowledge as to health and safety based upon our knowledge of the handling, storage and use of the product. It is not applicable to unusual or non-standard uses of the product or where instruction or recommendations are not followed.